

-----  
Midmark M11 - v1.0.0  
Total Cycles: 899

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2021 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
2:00 182.9 F 3.0  
4:00 185.9 F 6.8  
6:00 208.6 F 11.6  
8:00 222.2 F 6.6  
10:00 232.1 F 11.7  
12:00 236.7 F 8.9  
14:00 237.6 F 9.6  
16:00 248.2 F 14.1  
18:00 259.8 F 20.6  
20:00 270.5 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.1 F 29.1  
2:00 272.6 F 28.9  
3:00 271.7 F 28.7  
4:00 271.2 F 28.7  
  
Min 271.2 F 28.7  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 239.5 F 0.0  
10:00 226.0 F 0.0  
15:00 216.8 F 0.0  
20:00 205.2 F 0.0  
25:00 185.5 F 0.0  
30:00 183.6 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 21:20  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

06/15/2021 12:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_