
Midmark M11 - v1.0.0
Total Cycles: 1992

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 161.4 F 2.6
4:00 166.7 F 6.0
6:00 190.5 F 10.0
8:00 215.6 F 5.1
10:00 220.9 F 8.3
12:00 232.8 F 13.0
14:00 234.5 F 9.3
16:00 239.0 F 9.8
18:00 245.2 F 12.9
20:00 256.8 F 19.1
22:00 267.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.4 F 28.4
3:00 272.4 F 29.0
4:00 271.4 F 28.4

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 225.5 F 0.0
10:00 206.9 F 0.0
15:00 206.9 F 0.0
20:00 245.2 F 0.0
25:00 238.6 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

06/06/2022 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____