

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001541  
10:40 24/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:51  
STERILIZING 3:53  
136.0 C 221kPa 3:53  
136.3 C 224kPa 4:23  
136.4 C 226kPa 4:53  
136.4 C 225kPa 5:23  
136.4 C 227kPa 5:53  
136.2 C 224kPa 6:23  
136.4 C 226kPa 6:53  
136.4 C 224kPa 7:23  
136.4 C 224kPa 7:23  
VENTING 7:23  
AIR DRYING 8:17  
CYCLE COMPLETE 10:26  
-----

Digital Signature #  
81C14B0985C6D537  
-----