

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2626

Name: RITTER 1  
Sterilizer ID: V2178198  
11/30/2022 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 180.1 F 3.2  
4:00 185.0 F 7.5  
6:00 205.5 F 12.5  
8:00 218.3 F 6.4  
10:00 235.4 F 11.5  
12:00 233.1 F 8.4  
14:00 241.7 F 10.8  
16:00 245.6 F 13.3  
18:00 257.9 F 19.9  
20:00 269.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.6 F 29.3  
2:00 271.5 F 28.5  
3:00 272.3 F 29.4  
4:00 272.3 F 29.0  
  
Min 271.3 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 230.9 F 0.0  
10:00 240.0 F 0.0  
15:00 233.8 F 0.0  
20:00 226.1 F 0.0  
25:00 225.1 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 21:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

11/30/2022 10:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_