

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4000

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2023 11:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
2:00 176.6 F 3.0  
4:00 195.7 F 7.1  
6:00 212.4 F 11.8  
8:00 225.7 F 6.4  
10:00 234.8 F 11.3  
12:00 234.6 F 8.3  
14:00 241.1 F 11.4  
16:00 245.4 F 12.9  
18:00 257.0 F 19.3  
20:00 267.9 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.7 F 29.1  
2:00 271.3 F 28.4  
3:00 272.8 F 29.1  
4:00 271.2 F 28.4  
  
Min 271.2 F 28.4  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 231.9 F 0.0  
10:00 223.9 F 0.0  
15:00 225.5 F 0.0  
20:00 221.6 F 0.0  
25:00 221.6 F 0.0  
30:00 242.2 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 21:10  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:29  
CYCLE COMPLETE

11/09/2023 12:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_