
Midmark M11 - v1.0.0
Total Cycles: 1007

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2021 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.9 F 0.0
2:00 166.3 F 1.4
4:00 184.2 F 4.6
6:00 204.7 F 9.1
8:00 220.2 F 8.5
10:00 232.0 F 8.6
12:00 242.3 F 12.8
14:00 240.9 F 11.1
16:00 237.9 F 9.4
18:00 251.6 F 15.9
20:00 262.7 F 22.2
22:00 272.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.4 F 28.5
2:00 271.8 F 28.3
3:00 272.2 F 28.4
4:00 271.7 F 28.3

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 227.6 F 0.0
10:00 219.0 F 0.0
15:00 186.0 F 0.0
Dry Cycle Aborted

FILLING: 1:33
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:14
DRYING: 15:02
TOTAL CYCLE: 00:46:05

07/14/2021 08:42 AM
ACCEPTED: X
REJECTED:
Operator: U01