

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1505

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2022 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
2:00 185.1 F 3.5  
4:00 201.5 F 8.4  
6:00 219.1 F 7.6  
8:00 230.9 F 9.4  
10:00 236.5 F 8.7  
12:00 243.1 F 12.8  
14:00 242.5 F 11.5  
16:00 255.8 F 18.4  
18:00 267.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.6 F 29.0  
2:00 271.5 F 28.5  
3:00 272.2 F 28.7  
4:00 271.8 F 28.8  
  
Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 231.9 F 0.0  
10:00 226.8 F 0.0  
15:00 226.5 F 0.0  
20:00 226.2 F 0.0  
25:00 227.0 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

01/10/2022 02:30 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_