

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5443

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2025 01:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.6 F 0.0  
2:00 138.6 F 2.0  
4:00 147.7 F 4.8  
6:00 177.0 F 8.5  
8:00 202.0 F 12.6  
10:00 215.6 F 6.3  
12:00 229.3 F 11.2  
14:00 228.3 F 7.3  
16:00 239.9 F 13.0  
18:00 239.9 F 11.0  
20:00 252.4 F 17.3  
22:00 263.7 F 23.9  
24:00 271.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.6 F 29.6  
2:00 271.1 F 29.0  
3:00 272.7 F 29.8  
4:00 271.6 F 29.1

Min 271.0 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
5:00 214.7 F 0.0  
10:00 227.5 F 0.0  
15:00 231.4 F 0.0  
20:00 220.4 F 0.0  
25:00 223.5 F 0.0  
30:00 231.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:05  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:02:37  
CYCLE COMPLETE

03/25/2025 02:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_