

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2001

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2022 11:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
2:00 192.5 F 4.1  
4:00 223.9 F 9.5  
6:00 228.9 F 6.6  
8:00 234.8 F 11.0  
10:00 233.6 F 8.0  
12:00 241.0 F 10.7  
14:00 247.0 F 13.6  
16:00 259.9 F 20.7  
18:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.5  
1:00 272.5 F 29.0  
2:00 272.0 F 29.0  
3:00 272.0 F 28.5  
4:00 272.6 F 29.1  
  
Min 271.6 F 28.4  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 227.7 F 0.0  
10:00 211.1 F 0.0  
15:00 223.7 F 0.0  
20:00 232.4 F 0.0  
25:00 239.8 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

06/07/2022 12:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_