

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2071

Name: RITTER2  
Sterilizer ID: V2313228  
02/21/2023 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 180.1 F 2.2  
4:00 194.7 F 6.6  
6:00 214.9 F 12.3  
8:00 228.0 F 8.5  
10:00 237.9 F 10.8  
12:00 241.5 F 12.6  
14:00 241.7 F 11.9  
16:00 255.4 F 18.9  
18:00 267.2 F 26.3  
20:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.8  
1:00 272.9 F 30.4  
2:00 271.9 F 29.7  
3:00 272.9 F 30.6  
4:00 272.4 F 30.0

Min 271.8 F 29.7  
Max 272.9 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 218.3 F 0.0  
10:00 234.2 F 0.0  
15:00 232.8 F 0.0  
20:00 227.6 F 0.0  
25:00 229.4 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 20:10  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

02/21/2023 12:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_