
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001865
7:19 17/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:36
STERILIZING 3:42
136.2 C 221kPa 3:42
136.2 C 222kPa 4:12
136.3 C 223kPa 4:42
136.2 C 222kPa 5:12
136.0 C 221kPa 5:42
136.3 C 223kPa 6:12
136.4 C 224kPa 6:42
136.2 C 222kPa 7:12
136.2 C 222kPa 7:12
VENTING 7:12
AIR DRYING 8:03
CYCLE COMPLETE 8:49

Digital Signature #
0A5519B396CC8E33
