

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2034

Name: RITTER2  
Sterilizer ID: V2313228  
02/09/2023 03:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 174.6 F 2.8  
4:00 189.5 F 7.1  
6:00 213.9 F 12.5  
8:00 224.6 F 7.9  
10:00 237.8 F 12.4  
12:00 240.5 F 12.1  
14:00 241.1 F 11.5  
16:00 255.2 F 18.8  
18:00 267.5 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.7 F 29.8  
2:00 271.6 F 29.1  
3:00 272.7 F 30.1  
4:00 272.3 F 29.5  
  
Min 271.5 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 216.6 F 0.0  
10:00 234.6 F 0.0  
15:00 232.8 F 0.0  
20:00 224.5 F 0.0  
25:00 226.0 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 19:21  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:45  
CYCLE COMPLETE

02/09/2023 04:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_