
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001752
12:58 14/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:47
136.4 C 223kPa 3:47
136.2 C 223kPa 4:17
136.3 C 223kPa 4:47
136.4 C 224kPa 5:17
136.4 C 223kPa 5:47
136.4 C 222kPa 6:17
136.4 C 224kPa 6:47
136.4 C 222kPa 7:17
136.4 C 222kPa 7:17
VENTING 7:17
AIR DRYING 8:10
CYCLE COMPLETE 8:11

Digital Signature #
B3DD8BD4E5EB0E98
