

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2087

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2022 08:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.3 F 0.0  
2:00 171.1 F 2.1  
4:00 183.5 F 5.9  
6:00 202.7 F 10.8  
8:00 219.6 F 5.6  
10:00 231.3 F 10.8  
12:00 234.4 F 7.4  
14:00 243.6 F 12.4  
16:00 243.2 F 12.3  
18:00 255.9 F 18.9  
20:00 266.9 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.7 F 29.3  
2:00 271.7 F 28.8  
3:00 272.7 F 29.4  
4:00 271.7 F 28.7  
  
Min 271.6 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 201.8 F 0.0  
10:00 210.2 F 0.0  
15:00 213.7 F 0.0  
20:00 214.7 F 0.0  
25:00 214.7 F 0.0  
30:00 216.6 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 21:29  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:24  
CYCLE COMPLETE

07/05/2022 09:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_