
Midmark M11 - v1.0.5
Total Cycles: 325

Name: RITTER2
Sterilizer ID: V2313228
08/31/2021 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.7 F 0.0
2:00 115.4 F 1.1
4:00 121.6 F 3.7
6:00 155.4 F 7.8
8:00 185.0 F 12.9
10:00 202.1 F 6.9
12:00 233.0 F 12.2
14:00 235.3 F 9.7
16:00 233.2 F 9.1
18:00 247.7 F 14.4
20:00 259.2 F 20.9
22:00 269.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.1 F 30.1
2:00 272.6 F 29.8
3:00 272.2 F 30.2
4:00 272.4 F 29.9

Min 270.9 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 230.0 F 0.0
10:00 230.1 F 0.0
15:00 230.3 F 0.0
20:00 228.1 F 0.0
25:00 229.7 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:52
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

08/31/2021 12:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____