
Midmark M11 - v1.0.0
Total Cycles: 4783

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2024 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.4 F 0.0
2:00 110.2 F 1.1
4:00 140.3 F 3.6
6:00 161.8 F 6.4
8:00 181.1 F 9.6
10:00 199.8 F 13.5
12:00 208.3 F 4.7
14:00 220.0 F 8.2
16:00 231.9 F 12.4
18:00 232.4 F 8.2
20:00 243.3 F 13.3
22:00 241.9 F 11.4
24:00 253.7 F 17.4
26:00 264.4 F 23.7
28:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.8 F 29.2
2:00 272.0 F 28.5
3:00 272.2 F 29.0
4:00 272.7 F 29.1

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 199.3 F 0.0
10:00 205.2 F 0.0
15:00 212.8 F 0.0
20:00 224.9 F 0.0
25:00 229.5 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:39
CYCLE COMPLETE

08/01/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____