

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3149

Name: RITTER2  
Sterilizer ID: V2313228  
11/21/2023 09:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 71.8 F 0.0  
4:00 77.4 F 0.3  
6:00 103.4 F 1.7  
8:00 139.7 F 4.7  
10:00 167.1 F 8.5  
12:00 193.0 F 12.8  
14:00 211.6 F 5.6  
16:00 221.4 F 9.8  
18:00 229.9 F 9.4  
20:00 236.9 F 11.0  
22:00 235.1 F 8.8  
24:00 248.8 F 15.2  
26:00 260.2 F 21.6  
28:00 270.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.9 F 29.8  
2:00 271.5 F 28.9  
3:00 272.3 F 29.8  
4:00 272.3 F 29.4

Min 271.1 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 199.7 F 0.0  
10:00 193.1 F 0.0  
15:00 190.6 F 0.0  
20:00 188.5 F 0.0  
25:00 186.6 F 0.0  
30:00 185.3 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 28:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:06:29  
CYCLE COMPLETE

11/21/2023 10:57 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_