

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3787

Name: RITTER2  
Sterilizer ID: V2313228  
07/16/2024 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.1 F 0.0  
2:00 149.1 F 1.0  
4:00 149.1 F 3.7  
6:00 162.5 F 7.0  
8:00 184.3 F 10.6  
10:00 206.7 F 6.7  
12:00 218.5 F 8.3  
14:00 229.2 F 12.6  
16:00 230.9 F 9.1  
18:00 239.8 F 11.4  
20:00 242.2 F 12.6  
22:00 253.4 F 18.3  
24:00 263.6 F 24.4  
26:00 272.9 F 30.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.7  
1:00 271.9 F 30.4  
2:00 272.7 F 30.5  
3:00 272.1 F 30.1  
4:00 271.5 F 29.8

Min 271.0 F 29.6  
Max 272.8 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 213.9 F 0.0  
10:00 227.2 F 0.0  
15:00 230.0 F 0.0  
20:00 222.1 F 0.0  
25:00 220.6 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:25  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

07/16/2024 11:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_