
Midmark M11 - v1.0.0
Total Cycles: 4302

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2024 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.6 F 0.0
2:00 90.3 F 0.6
4:00 110.5 F 2.3
6:00 138.2 F 4.9
8:00 165.4 F 7.7
10:00 190.4 F 11.5
12:00 206.1 F 2.5
14:00 220.1 F 5.6
16:00 227.1 F 8.9
18:00 235.9 F 12.9
20:00 233.8 F 8.8
22:00 244.0 F 12.7
24:00 242.8 F 11.7
26:00 253.8 F 17.3
28:00 263.8 F 23.3
30:00 272.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.0 F 28.8
2:00 272.5 F 28.8
3:00 271.6 F 28.2
4:00 272.1 F 28.9

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 216.3 F 0.0
10:00 226.3 F 0.0
15:00 230.5 F 0.0
20:00 227.8 F 0.0
25:00 230.4 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

02/14/2024 03:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____