
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000474
13:58 12/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:19
STERILIZING 4:33
136.0 C 228kPa 4:33
136.4 C 233kPa 5:03
136.3 C 233kPa 5:33
136.2 C 230kPa 6:03
136.4 C 233kPa 6:33
136.3 C 233kPa 7:03
136.2 C 231kPa 7:33
136.4 C 232kPa 8:03
136.4 C 232kPa 8:03
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 24:00

Digital Signature #
B27E33F9E1EF382C
