

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1335

Name: RITTER 1  
Sterilizer ID: V2178198  
11/03/2021 10:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
2:00 179.0 F 2.9  
4:00 191.2 F 7.6  
6:00 212.7 F 13.4  
8:00 225.6 F 9.2  
10:00 232.8 F 8.9  
12:00 244.9 F 12.7  
14:00 241.9 F 11.0  
16:00 255.0 F 17.7  
18:00 266.1 F 24.6  
20:00 273.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.9 F 29.3  
2:00 271.2 F 28.8  
3:00 272.8 F 29.1  
4:00 271.6 F 29.1  
  
Min 271.1 F 28.6  
Max 273.2 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 227.3 F 0.0  
10:00 224.0 F 0.0  
15:00 224.2 F 0.0  
20:00 224.0 F 0.0  
25:00 172.2 F 0.0  
30:00 165.0 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

11/03/2021 11:41 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01