

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001618  
11:28 11/25/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:31  
PRESSURIZING 6:17  
STERILIZING 7:52  
136.0 C 220kPa 7:52  
136.4 C 223kPa 8:22  
136.4 C 224kPa 8:52  
136.4 C 223kPa 9:22  
136.0 C 222kPa 9:52  
136.5 C 221kPa 10:22  
136.4 C 223kPa 10:52  
136.2 C 218kPa 11:22  
VENTING 11:22  
AIR DRYING 12:31  
CYCLE COMPLETE 34:03

-----  
Digital Signature #  
AB14B16595848CF4  
-----