

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4769

Name: RITTER 1  
Sterilizer ID: V2178198  
07/26/2024 11:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.4 F 0.0  
2:00 167.9 F 3.0  
4:00 172.9 F 6.6  
6:00 192.6 F 10.6  
8:00 211.3 F 5.0  
10:00 222.6 F 8.1  
12:00 233.1 F 12.7  
14:00 232.1 F 9.0  
16:00 239.3 F 10.7  
18:00 244.5 F 12.7  
20:00 255.8 F 18.6  
22:00 266.4 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.1  
2:00 271.4 F 28.7  
3:00 272.4 F 29.2  
4:00 271.5 F 28.6

Min 271.3 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 221.9 F 0.0  
10:00 227.6 F 0.0  
15:00 236.7 F 0.0  
20:00 227.4 F 0.0  
25:00 224.4 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 23:41  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:02:21  
CYCLE COMPLETE

07/26/2024 12:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_