

-----  
Midmark M11 - v1.0.0  
Total Cycles: 942

Name: RITTER 1  
Sterilizer ID: V2178198  
06/25/2021 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.0 F 0.0  
2:00 120.6 F 0.0  
4:00 137.6 F 0.2  
6:00 162.2 F 2.4  
8:00 185.7 F 5.2  
10:00 202.7 F 8.5  
12:00 216.1 F 12.1  
14:00 222.5 F 6.3  
16:00 232.7 F 10.9  
18:00 233.1 F 7.8  
20:00 243.8 F 12.8  
22:00 241.8 F 11.1  
24:00 253.2 F 16.8  
26:00 263.1 F 22.7  
28:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.0 F 28.9  
2:00 272.5 F 28.8  
3:00 271.8 F 28.3  
4:00 271.2 F 28.1

Min 271.2 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
5:00 219.9 F 0.0  
10:00 225.2 F 0.0  
15:00 213.7 F 0.0  
20:00 215.1 F 0.0  
25:00 228.9 F 0.0  
30:00 232.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:28  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:06:49  
CYCLE COMPLETE

06/25/2021 10:13 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_