
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000103
13:27 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:27
PRESSURIZING 4:28
STERILIZING 6:02
136.5 C 224kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 7:02
136.4 C 227kPa 7:32
136.4 C 225kPa 8:02
136.4 C 224kPa 8:32
136.3 C 225kPa 9:02
136.2 C 224kPa 9:32
136.2 C 224kPa 9:32
VENTING 9:32
AIR DRYING 10:32
CYCLE COMPLETE 25:14

Digital Signature #
29D966ABDF4FCE2A
