
Midmark M11 - v1.0.5
Total Cycles: 4451

Name: RITTER2
Sterilizer ID: V2313228
03/25/2025 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 70.7 F 0.1
4:00 72.5 F 0.8
6:00 92.0 F 2.6
8:00 121.8 F 5.2
10:00 165.2 F 7.9
12:00 165.0 F 11.2
14:00 208.4 F 4.3
16:00 211.2 F 5.8
18:00 218.9 F 8.9
20:00 229.2 F 12.5
22:00 230.5 F 8.4
24:00 241.0 F 13.0
26:00 239.6 F 11.0
28:00 250.4 F 16.4
30:00 260.0 F 22.0
32:00 269.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.9
1:00 271.6 F 29.9
2:00 271.6 F 30.0
3:00 271.7 F 29.9
4:00 271.7 F 29.8

Min 271.5 F 29.8
Max 272.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 224.1 F 0.0
10:00 208.2 F 0.0
15:00 187.1 F 0.0
20:00 182.3 F 0.0
25:00 192.1 F 0.0
30:00 196.4 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:11:44
CYCLE COMPLETE

03/25/2025 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____