

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4227

Name: RITTER2  
Sterilizer ID: V2313228  
12/18/2024 03:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.1 F 0.0  
2:00 135.4 F 1.7  
4:00 150.6 F 4.5  
6:00 175.5 F 8.4  
8:00 194.6 F 12.7  
10:00 216.9 F 6.2  
12:00 228.8 F 10.7  
14:00 231.8 F 7.5  
16:00 241.3 F 12.7  
18:00 240.1 F 11.1  
20:00 252.1 F 16.9  
22:00 262.7 F 23.1  
24:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 272.1 F 29.6  
2:00 272.1 F 29.6  
3:00 272.6 F 29.8  
4:00 272.6 F 30.0

Min 271.8 F 29.6  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 222.1 F 0.0  
10:00 223.5 F 0.0  
15:00 231.1 F 0.0  
20:00 226.4 F 0.0  
25:00 228.9 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 25:10  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:04:02  
CYCLE COMPLETE

12/18/2024 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_