

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002519  
15:25 19/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:41  
STERILIZING 3:42  
136.4 C 224kPa 3:42  
136.4 C 225kPa 4:12  
136.4 C 226kPa 4:42  
136.2 C 225kPa 5:12  
136.3 C 225kPa 5:42  
136.4 C 225kPa 6:12  
136.4 C 226kPa 6:42  
136.4 C 225kPa 7:12  
136.4 C 225kPa 7:12  
VENTING 7:12  
AIR DRYING 8:08  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
A092D29EB84EBC39  
-----