

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4535

Name: RITTER 1  
Sterilizer ID: V2178198  
05/08/2024 01:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.6 F 0.0  
2:00 94.9 F 0.1  
4:00 120.5 F 1.7  
6:00 153.4 F 4.8  
8:00 190.5 F 8.0  
10:00 205.3 F 12.5  
12:00 215.6 F 4.6  
14:00 223.4 F 8.6  
16:00 235.3 F 12.8  
18:00 234.7 F 9.0  
20:00 241.0 F 10.5  
22:00 244.5 F 12.7  
24:00 256.6 F 19.0  
26:00 267.5 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.7 F 29.2  
2:00 271.5 F 28.3  
3:00 272.5 F 29.3  
4:00 272.3 F 28.8

Min 271.3 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 215.3 F 0.0  
10:00 207.1 F 0.0  
15:00 239.9 F 0.0  
20:00 231.9 F 0.0  
25:00 228.6 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 27:27  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:05:41  
CYCLE COMPLETE

05/08/2024 02:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_