
Midmark M11 - v1.0.5
Total Cycles: 4161

Name: RITTER2
Sterilizer ID: V2313228
11/27/2024 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.3 F 0.0
2:00 118.0 F 0.0
4:00 117.0 F 0.1
6:00 152.3 F 1.8
8:00 172.4 F 4.9
10:00 193.2 F 8.5
12:00 209.0 F 12.5
14:00 220.7 F 6.7
16:00 235.9 F 11.9
18:00 236.0 F 9.9
20:00 237.0 F 9.2
22:00 249.2 F 15.5
24:00 260.9 F 22.1
26:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.7 F 29.9
2:00 271.8 F 29.2
3:00 271.5 F 29.6
4:00 272.8 F 29.7

Min 271.1 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 226.5 F 0.0
10:00 229.7 F 0.0
15:00 203.4 F 0.0
20:00 195.4 F 0.0
25:00 192.4 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:40
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

11/27/2024 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____