
Midmark M11 - v1.0.0
Total Cycles: 3100

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.9 F 0.0
2:00 197.1 F 4.1
4:00 214.5 F 9.7
6:00 225.9 F 6.1
8:00 235.6 F 11.2
10:00 235.3 F 8.7
12:00 239.4 F 9.8
14:00 248.3 F 14.4
16:00 261.0 F 21.5
18:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.4 F 28.8
2:00 271.9 F 28.7
3:00 272.4 F 28.8
4:00 271.5 F 28.4

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 225.2 F 0.0
10:00 241.4 F 0.0
15:00 234.8 F 0.0
20:00 227.4 F 0.0
25:00 229.1 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

04/12/2023 12:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____