
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001698
10:59 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:53
STERILIZING 3:55
136.0 C 220kPa 3:55
136.4 C 224kPa 4:25
136.0 C 221kPa 4:55
136.2 C 223kPa 5:25
136.3 C 223kPa 5:55
136.3 C 221kPa 6:25
136.4 C 225kPa 6:55
136.4 C 224kPa 7:25
136.4 C 224kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 8:20

Digital Signature #
37EB7677A8FC0BEC
