
Midmark M11 - v1.0.5
Total Cycles: 2322

Name: RITTER2
Sterilizer ID: V2313228
05/02/2023 07:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.6 F 0.0
2:00 78.1 F 0.0
4:00 81.2 F 0.0
6:00 99.9 F 0.0
8:00 144.9 F 2.6
10:00 179.6 F 5.9
12:00 190.2 F 10.0
14:00 217.7 F 8.9
16:00 231.0 F 7.5
18:00 241.5 F 12.0
20:00 235.6 F 8.9
22:00 239.5 F 10.9
24:00 245.5 F 13.4
26:00 257.3 F 19.7
28:00 268.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.4
2:00 271.1 F 28.9
3:00 272.5 F 29.5
4:00 271.3 F 28.7

Min 271.1 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 204.3 F 0.0
10:00 193.5 F 0.0
15:00 196.2 F 0.0
20:00 194.0 F 0.0
25:00 193.2 F 0.0
30:00 194.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:18
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:07:21
CYCLE COMPLETE

05/02/2023 08:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____