
Midmark M11 - v1.0.0
Total Cycles: 881

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 09:58 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.0 F 0.0
2:00 159.1 F 1.8
4:00 170.5 F 4.6
6:00 193.3 F 8.5
8:00 210.1 F 13.1
10:00 222.5 F 6.2
12:00 234.8 F 10.8
14:00 232.4 F 7.5
16:00 244.3 F 13.0
18:00 242.8 F 11.5
20:00 254.9 F 17.6
22:00 265.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.7
2:00 271.3 F 27.9
3:00 272.5 F 28.8

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 214.1 F 0.0
10:00 200.1 F 0.0
15:00 216.8 F 0.0
20:00 207.8 F 0.0
25:00 233.5 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:52
STERILIZING: 3:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:35
CYCLE COMPLETE

06/10/2021 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____