

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2098

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2023 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.8 F 0.0  
2:00 79.5 F 0.0  
4:00 88.5 F 0.0  
6:00 110.9 F 0.2  
8:00 168.8 F 0.3  
10:00 195.9 F 2.1  
12:00 205.4 F 6.0  
14:00 215.5 F 9.8  
16:00 228.4 F 10.9  
18:00 232.9 F 9.4  
20:00 238.5 F 10.8  
22:00 243.0 F 12.6  
24:00 241.9 F 11.6  
26:00 254.6 F 18.2  
28:00 265.9 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.4 F 29.3  
2:00 271.9 F 29.4  
3:00 272.1 F 29.1  
4:00 272.1 F 29.6

Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 221.5 F 0.0  
10:00 194.2 F 0.0  
15:00 192.6 F 0.0  
20:00 192.3 F 0.0  
25:00 191.7 F 0.0  
30:00 202.7 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 29:38  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:07:41  
CYCLE COMPLETE

02/28/2023 09:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_