
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001819
8:08 10/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:59
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:59

Digital Signature #
38C7A18BD211B8FB
