
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002290
9:52 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:08
STERILIZING 3:08
136.0 C 220kPa 3:08
136.4 C 225kPa 3:38
136.2 C 223kPa 4:08
136.4 C 225kPa 4:38
136.4 C 224kPa 5:08
136.2 C 223kPa 5:38
136.4 C 224kPa 6:08
136.3 C 224kPa 6:38
136.3 C 224kPa 6:38
VENTING 6:38
AIR DRYING 7:29
CYCLE COMPLETE 37:29

Digital Signature #
DD615A9FF82C3DB5
