
Midmark M11 - v1.0.0
Total Cycles: 3923

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2023 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.6 F 0.0
2:00 193.6 F 3.2
4:00 191.7 F 8.1
6:00 214.4 F 11.3
8:00 226.2 F 7.8
10:00 237.8 F 13.0
12:00 239.1 F 10.3
14:00 235.6 F 8.0
16:00 249.9 F 15.0
18:00 261.7 F 21.7
20:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.1
3:00 271.9 F 28.8
4:00 272.6 F 28.7

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 221.4 F 0.0
10:00 200.5 F 0.0
15:00 217.4 F 0.0
20:00 202.1 F 0.0
25:00 211.0 F 0.0
30:00 239.9 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

10/19/2023 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____