
Midmark M11 - v1.0.0
Total Cycles: 4653

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2024 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 178.1 F 3.9
4:00 181.0 F 8.1
6:00 207.4 F 12.8
8:00 217.6 F 6.3
10:00 231.8 F 11.3
12:00 231.9 F 7.7
14:00 244.2 F 13.6
16:00 243.3 F 11.9
18:00 255.8 F 18.4
20:00 266.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.4 F 29.1
4:00 272.3 F 28.7

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 213.0 F 0.0
10:00 208.6 F 0.0
15:00 204.8 F 0.0
20:00 227.9 F 0.0
25:00 222.2 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

06/18/2024 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____