
Midmark M11 - v1.0.5
Total Cycles: 3270

Name: RITTER2
Sterilizer ID: V2313228
01/03/2024 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.8 F 0.0
2:00 127.7 F 0.5
4:00 135.8 F 2.6
6:00 160.6 F 6.0
8:00 185.1 F 10.0
10:00 206.4 F 8.1
12:00 219.6 F 7.6
14:00 232.5 F 12.2
16:00 231.8 F 9.0
18:00 240.3 F 11.0
20:00 243.7 F 13.0
22:00 254.6 F 19.0
24:00 265.4 F 25.4
26:00 271.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.7
1:00 272.3 F 30.4
2:00 273.3 F 30.5
3:00 272.6 F 30.0
4:00 271.7 F 29.6

Min 271.4 F 29.5
Max 273.5 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 231.6 F 0.0
10:00 219.6 F 0.0
15:00 203.4 F 0.0
20:00 202.4 F 0.0
25:00 205.5 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:02
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

01/03/2024 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____