
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001005
10:07 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:50
136.4 C 232kPa 3:50
136.3 C 232kPa 4:20
136.3 C 232kPa 4:50
136.4 C 233kPa 5:20
136.2 C 229kPa 5:50
136.4 C 233kPa 6:20
136.4 C 233kPa 6:50
136.4 C 233kPa 7:20
136.4 C 233kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 15:58

Digital Signature #
243D031C4180EC45
