
Midmark M11 - v1.0.5
Total Cycles: 755

Name: RITTER2
Sterilizer ID: V2313228
02/02/2022 05:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.3 F 0.0
2:00 131.9 F 1.7
4:00 155.5 F 4.9
6:00 181.9 F 9.4
8:00 209.0 F 7.2
10:00 221.4 F 8.6
12:00 231.9 F 11.2
14:00 237.7 F 11.4
16:00 237.9 F 10.0
18:00 251.7 F 16.8
20:00 263.7 F 23.9
22:00 271.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.6 F 29.8
2:00 271.4 F 29.0
3:00 271.8 F 29.8
4:00 272.3 F 29.7

Min 271.1 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 211.8 F 0.0
10:00 220.0 F 0.0
15:00 224.4 F 0.0
20:00 224.0 F 0.0
25:00 228.0 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

02/02/2022 06:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____