
Midmark M11 - v1.0.0
Total Cycles: 2398

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2022 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 200.3 F 3.6
4:00 204.3 F 7.7
6:00 226.0 F 13.1
8:00 229.2 F 7.8
10:00 240.8 F 13.2
12:00 239.1 F 10.6
14:00 237.0 F 8.7
16:00 251.0 F 15.7
18:00 262.8 F 22.5
20:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.7 F 28.6
3:00 272.4 F 28.9
4:00 271.5 F 28.4

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 231.1 F 0.0
10:00 245.5 F 0.0
15:00 240.3 F 0.0
20:00 228.7 F 0.0
25:00 223.2 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

10/04/2022 03:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____