
North York, ON
Canada

STATIM 2000 S203R750

SN 500919I00007

Unit # : 000

WATER QUALITY

0.5uS / 0.3ppm

CYCLE NUMBER 001903

12:12 26/07/2024

UNWRAPPED

135 C for 3.5 min.

CYCLE START 0:00

CONDITIONING 1:53

PRESSURIZING 1:56

STERILIZING 3:00

136.4 C 227kPa 3:00

136.4 C 227kPa 3:30

136.3 C 226kPa 4:00

136.4 C 226kPa 4:30

136.4 C 226kPa 5:00

136.4 C 227kPa 5:30

136.2 C 223kPa 6:00

136.4 C 225kPa 6:30

136.4 C 225kPa 6:30

VENTING 6:30

AIR DRYING 7:23

CYCLE COMPLETE 37:23

Digital Signature #

3A4F8C321B43C484
