
Midmark M11 - v1.0.5
Total Cycles: 2751

Name: RITTER2
Sterilizer ID: V2313228
08/10/2023 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.4 F 0.0
2:00 193.9 F 3.6
4:00 201.7 F 9.2
6:00 219.8 F 7.4
8:00 230.9 F 10.5
10:00 232.3 F 7.8
12:00 240.5 F 10.9
14:00 246.6 F 14.2
16:00 259.2 F 21.2
18:00 270.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.4 F 29.4
2:00 272.5 F 30.0
3:00 271.9 F 29.1
4:00 272.7 F 29.8

Min 271.7 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 209.5 F 0.0
10:00 220.0 F 0.0
15:00 222.4 F 0.0
20:00 221.3 F 0.0
25:00 222.0 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

08/10/2023 12:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____