
Midmark M11 - v1.0.0
Total Cycles: 1968

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2022 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 198.2 F 4.6
4:00 206.7 F 9.8
6:00 225.4 F 6.1
8:00 234.8 F 10.4
10:00 236.0 F 7.4
12:00 243.2 F 11.7
14:00 245.2 F 13.0
16:00 258.4 F 20.0
18:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.2 F 28.4
3:00 272.5 F 29.1
4:00 271.5 F 28.3

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 219.1 F 0.0
10:00 232.7 F 0.0
15:00 225.5 F 0.0
20:00 221.7 F 0.0
25:00 222.5 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

05/27/2022 04:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____