
Midmark M11 - v1.0.0
Total Cycles: 1921

Name: RITTER 1
Sterilizer ID: V2178198
05/09/2022 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 155.6 F 2.2
4:00 176.5 F 5.8
6:00 198.0 F 10.3
8:00 219.8 F 4.9
10:00 225.3 F 8.1
12:00 238.3 F 12.9
14:00 237.4 F 9.6
16:00 237.3 F 9.0
18:00 247.3 F 13.6
20:00 259.1 F 20.0
22:00 269.8 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 273.0 F 29.1
2:00 272.2 F 28.5
3:00 271.9 F 28.7
4:00 273.1 F 29.0

Min 271.4 F 28.2
Max 273.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 200.4 F 0.0
10:00 198.0 F 0.0
15:00 197.3 F 0.0
20:00 196.1 F 0.0
25:00 194.1 F 0.0
30:00 199.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

05/09/2022 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____