
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000439
11:00 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:36
136.4 C 231kPa 3:36
136.3 C 232kPa 4:06
136.4 C 231kPa 4:36
136.3 C 232kPa 5:06
136.4 C 232kPa 5:36
136.3 C 233kPa 6:06
136.3 C 232kPa 6:36
136.4 C 231kPa 7:06
136.4 C 231kPa 7:06
VENTING 7:06
AIR DRYING 8:00
CYCLE COMPLETE 8:36

Digital Signature #
BD1A67AC90B97A6B
