
Midmark M11 - v1.0.0
Total Cycles: 3301

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2023 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.1 F 0.0
2:00 78.9 F 0.0
4:00 107.9 F 0.0
6:00 111.0 F 0.0
8:00 147.4 F 0.0
10:00 179.1 F 1.7
12:00 194.3 F 5.0
14:00 206.2 F 8.0
16:00 217.7 F 11.3
18:00 224.6 F 5.6
20:00 232.6 F 9.4
22:00 240.3 F 11.5
24:00 240.9 F 11.0
26:00 236.5 F 8.7
28:00 250.2 F 15.0
30:00 260.9 F 21.0
32:00 270.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.5 F 28.7
2:00 271.6 F 28.0
3:00 271.9 F 28.5
4:00 272.4 F 28.5

Min 271.2 F 27.8
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 212.3 F 0.0
10:00 219.6 F 0.0
15:00 223.0 F 0.0
20:00 224.5 F 0.0
25:00 231.2 F 0.0
30:00 219.1 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:49
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:10:52
CYCLE COMPLETE

05/30/2023 08:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____