
Midmark M11 - v1.0.5
Total Cycles: 2026

Name: RITTER2
Sterilizer ID: V2313228
02/08/2023 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.4 F 0.0
2:00 144.4 F 0.0
4:00 167.8 F 1.2
6:00 192.4 F 5.5
8:00 216.6 F 10.3
10:00 224.5 F 6.6
12:00 235.7 F 11.3
14:00 235.7 F 9.3
16:00 237.7 F 9.4
18:00 249.9 F 15.6
20:00 262.4 F 22.8
22:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.6
2:00 271.8 F 28.8
3:00 272.5 F 29.8
4:00 272.4 F 29.3

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 198.6 F 0.0
10:00 221.7 F 0.0
15:00 221.2 F 0.0
20:00 221.7 F 0.0
25:00 221.9 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

02/08/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____